INCUBATOR YP90A DAVID

Infant Incubator



registration certificate number: 20143541903














- . Two control modes: Air mode and Baby mode controlled by microcomputer;
- Set temperature, Air temperature, Baby temperature are displayed separately;
- Passive humidification;
- · Inclination of bassinet is adjustable;
- Multiple failure alarm indications;
- · RS-232 connector for data output;
- · Oxygen input interface;
- · Low noise DC motor.

